

## **BLOT-LINE Chlamydia pneumoniae IgG**



EAN Code: 8595635304415 Catalog number: CpGL20 Package size: 20 tests

Storage: 2-8 °C

**Producer:** TestLine Clinical Diagnostics s.r.o.

## **Description:**

- Antigen used: combination of selected parts of the specific antigens of Ch. pneumoniae (MOMP, MOMP1, OMP2, OMP4, OMP5, P54).
- Recombinant antigens are transferred to the nitrocellulose membrane (i.e. to the respective

BLOT-LINE (BL) strips).

- If specific antibodies are present in the sample, they will bind to the respective antigens.
- The complex is labelled with Conjugate and detected through a colour reaction with substrate (BCIP/NBT).
- The kit enables 20 tests.
- Short incubation periods, total assay time: approximately 1.5 h.
- High sensitivity and specificity.
- All reagents supplied are ready to use.
- The kit may be used sequentially for smaller batches of samples (reagents are provided in sufficient quantities).
- Positive and Negative controls included enabling validation of the tests.
- BL strips equipped with conjugate control band and a control band indicating kits functionality and sensitivity.
- Colour identification in compliance with antibody classes.
- Identical (i.e. interchangeable) reagents (except for Conjugate) in all BLOT-LINE kits.
- Evaluation of results the intensity of the bands can be evaluated visually or by means of

Stránka 1 / 2 Exportováno 8. 12. 2025



software (Immunoblot Software).

## **Applications:**

- Laboratory test for the detection of chlamydia infection.
- Confirmatory test to ELISA.

## Brief assay procedure:

- 1. Dilute serum/plasma samples (1:51).
- 2. Pipette diluted samples and controls into the channels of the tray and insert the BL strips.
- Incubate for 30 min at room temperature using a shaking apparatus. 3.
- 4. Wash 3 times for 5 min.
- 5. Add Conjugate.
- Incubate for 30 min at room temperature using a shaking apparatus. 6.
- 7. Wash 3 times for 5 min.
- 8. Add substrate (BCIP/NBT).
- Incubate for 15 min at room temperature using a shaking apparatus.
- 10. Wash 2 times for 5 min in distilled water.
- 11. Dry the strips and evaluate results using the validation strip enclosed or using Immunoblot Software.

Exportováno 8. 12. 2025 Stránka 2 / 2