

## BlueBLOT-LINE Borrelia IgG



**EAN Code:** 8595635305214

**Catalog number:** BD-BGL024

**Package size:** 24 tests

**Storage:** 2-8 °C

**Producer:** TestLine Clinical Diagnostics s.r.o.

### Description:

- Each single test contains all the reagents (ready-to-use) necessary for performing the test.
- Antigen used: combination of selected parts of the specific antigens of *Borrelia* sp. (VlsE Ba, VlsE Bg, VlsE Bs, p83, p58, p41, p39, OspB, OspA Ba, OspA Bg, OspA Bs, OspC, p17, NapA), highly specific antigens of *Anaplasma phagocytophilum* (p44, OmpA) and specific antigen of *Treponema pallidum* TpN17.
- Recombinant antigens are transferred to the nitrocellulose membrane, if specific antibodies are present in the sample, they will bind to the respective antigens. The complex is labelled with Conjugate and detected through a colour reaction with substrate.
- Short incubation periods, total assay time: approximately 1.25 h.
- Evaluation using Immunoblot Software (TestLine).

### Applications:

- Diagnostics of Lyme borreliosis; confirmation for ELISA tests.
- Diagnostics of Anaplasmosis.

### Brief assay procedure:

1. Insert testing strips and cartridges into the BlueDiver Instrument.
2. Pipette samples.
3. Fully automated incubation and washing steps in BlueDiver Instrument
4. Incubation 30 minutes.
5. Wash 3 x 2 min.
6. Conjugate dispensing.
7. Incubation 20 min.
8. Wash 3 x 2 min.
9. Substrate dispensing.
10. Incubation 10 min.
11. Wash 1x 2 min.
12. Evaluation of test results using the BlueScan scanner and Immunoblot Software.

