

BlueBLOT-LINE Borrelia IgG



EAN Code: 8595635305214 **Catalog number**: BD-BGL024

Package size: 24 tests

Storage: 2-8 °C

Producer: TestLine Clinical Diagnostics s.r.o.

Description:

- Each single test contains all the reagents (ready-to-use) necessary for performing the test.
- Antigen used: Antigen used: combination of selected parts of the specific antigens of Borrelia sp. (VISE Ba, VISE Bg, VISE Bs ,p83, p58, p41, p39, OspB, OspA Ba, OspA Bg, OspA Bs, OspC, p17, NapA), highly specific antigens of Anaplasma phagocytophilum (p44, OmpA) and specific antigen of Treponema pallidum TpN17.
- Recombinant antigens are transferred to the nitrocellulose membrane, if specific antibodies are
 present in the sample, they will bind to the respective antigens. The complex is labelled with
 Conjugate and detected through a colour reaction with substrate.
- Short incubation periods, total assay time: approximately 1.25 h.
- Evaluation using Immunoblot Software (TestLine).

Applications:

- Diagnostics of Lyme borreliosis; confirmation for ELISA tests.
- Diagnostics of Anaplasmosis.

Brief assay procedure:

- 1. Insert testing strips and cartridges into the BlueDiver Instrument.
- 2. Pipette samples.
- 3. Fully automated incubation and washing steps in BlueDiver Instrument
- 4. Incubation 30 minutes.
- 5. Wash 3 x 2 min.
- 6. Conjugate dispensing.
- 7. Incubation 20 min.
- 8. Wash 3 x 2 min.
- 9. Substrate dispensing.
- 10. Incubation 10 min.
- 11. Wash 1x 2 min.
- 12. Evaluation of test results using the BlueScan scanner and Immunoblot Software.

Exportováno 15. 12. 2025 Stránka 1 / 2





Exportováno 15. 12. 2025 Stránka 2 / 2