

Control set CLIA U1RNP

EAN Code: 8595635311055 Catalog number: CL-RNPCON Package size: 2x20 tests

Storage: 2-8 °C

Producer: TestLine Clinical Diagnostics s.r.o.



- The kit is intended for use together with CLIA kits on the KleeYa® analyzer.
- The kit includes positive and negative control sera, which are used to verify the functionality of CLIA tests and the accuracy of obtained results.

For more information, please refer to BiovendorGroup CLIA webpage...

Exportováno 5. 12. 2025 Stránka 1 / 1