


Client report

	No of participants	No of responded participants	Response percentage
Bordetella pertussis, antibodies, April, 2-2025	77	70	90.9 %


Summary Result 1



Summary	Own score	Max score	Own success rate	Difference	AVR success rate
Sample S001	10	10	100 %	5.7 %	94.3 %
Sample S002	10	10	100 %	13.2 %	86.8 %
Average:			100 %	9.5 %	90.5 %


History	Test nr.	Own success rate	Difference	AVR success rate
Round 2024-4	1-1	87.5 %	0.3 %	87.2 %
Round 2024-4	1-2	87.5 %	0.3 %	87.2 %
Round 2024-4	1-3	87.5 %	0.3 %	87.2 %
Round 2024-2	1-1	100 %	10.6 %	89.4 %
Round 2024-2	1-2	100 %	10.6 %	89.4 %
Round 2024-2	1-3	100 %	10.6 %	89.4 %
Round 2023-4	1-1	60 %	-25.2 %	85.2 %
Round 2023-4	1-2	60 %	-25.2 %	85.2 %

Sample S001 Result 1




Sample S001 results	Responded	Own score	Max score	Own success rate	Difference	AVR success rate	Count
	Clinical interpretation	4	4	100 %	10.9 %	89.1 %	74
	BopeAbA	2	2	100 %	6.1 %	93.9 %	49
	BopeAbG	2	2	100 %	0 %	100 %	74
	BopeAbM	2	2	100 %	0 %	100 %	12
Total:		10	10	100 %	5.7 %	94.3 %	209

Sample S001 Clinical interpretation




Clinical interpretation	Interpretation	Interpretation count	Own score	Max score	Own success rate	Difference	AVR success rate
	☒ Positive, sign of infection	41	4	4	100 %	0 %	100 %
	Longstanding infection, serological scar	5	-				0 %
	Laboratory does not give clinical interpretations	28	-				-
	Total:	74	4	4	100 %	10.9 %	89.1 %


Sample S001 BopeAbA



Sample S001 BopeAbG



Sample S001 BopeAbM



BopeAbA	Interpretation	Method	Interpretation count	Method count	Own score	Max score	Own success rate	Difference	AVR success rate
	Negative		2		-				0 %
		Euroimmun Anti- B. pertussis Toxin ELISA IgA		1					
		Euroimmun Euroline Bordetella pertussis IgA		1					
	☒ Positive		46		2	2	100 %	0 %	100 %
		Diasorin LIAISON® Bordetella pertussis Toxin IgA		12					
		Euroimmun Anti- B. pertussis Toxin ELISA IgA		11					
		Euroimmun Anti-Bordetella FHA ELISA IgA		3					
		Euroimmun Euroline Bordetella pertussis IgA		2					
		Novatec NovaLisa B. pertussis IgA ELISA		1					
		Orgentec B. pertussis TOXIN IgA ELISA		12					
		TestLine BLOT-LINE Bordetella IgA		1					
		☒ TestLine EIA Bordetella pertussis Toxin IgA		2					
		TestLine Microblot-Array Bordetella IgA		1					
		Virion/Serion B. pertussis Toxin IgA		1					
	Uncertain		1		-				0 %
		Euroimmun Anti- B. pertussis Toxin ELISA IgA		1					
	Total:		49		2	2	100 %	6.1 %	93.9 %

BopeAbG	Interpretation	Method	Interpretation count	Method count	Own score	Max score	Own success rate	Difference	AVR success rate
	☒ Positive		74		2	2	100 %	0 %	100 %
		Diasorin LIAISON® Bordetella pertussis Toxin IgG		25					
		Euroimmun Anti-B. pertussis Toxin ELISA IgG		17					
		Euroimmun Anti-Bordetella FHA ELISA IgG		3					
		Euroimmun Euroline Bordetella pertussis IgG		3					
		Novagnost Bordetella pertussis toxin (PT) IgG		1					
		Novatec NovaLisa B. pertussis IgG ELISA		2					
		Orgentec B. pertussis TOXIN IgG ELISA		12					
		TestLine BLOT-LINE Bordetella IgG		1					
		☒ TestLine EIA Bordetella pertussis Toxin IgG		2					
		TestLine Microblot-Array Bordetella IgG		1					

		Vircell VIRCLIA Monotest Bordetella Pertussis Toxin IgG		4					
		Virion/Serion B. pertussis Toxin IgG		3					
	Total:		74		2	2	100 %	0 %	100 %


BopeAbM	Interpretation	Method	Interpretation count	Method count	Own score	Max score	Own success rate	Difference	AVR success rate
	☒ Negative		12		2	2	100 %	0 %	100 %
		Euroimmun Anti-B. pertussis ELISA IgM		9					
		Novatec NovaLisa B. pertussis IgM ELISA		1					
		☒ TestLine EIA Bordetella pertussis Toxin IgM		2					
	Total:		12		2	2	100 %	0 %	100 %

Sample S002 Result 1




Sample S002 results	Responded	Own score	Max score	Own success rate	Difference	AVR success rate	Count
	Clinical interpretation	4	4	100 %	19.6 %	80.4 %	74
	BopeAbA	2	2	100 %	14.3 %	85.7 %	49
	BopeAbG	2	2	100 %	5.4 %	94.6 %	74
	BopeAbM	2	2	100 %	8.3 %	91.7 %	12
	Total:	10	10	100 %	13.2 %	86.8 %	209

Sample S002 Clinical interpretation




Clinical interpretation	Interpretation	Interpretation count	Own score	Max score	Own success rate	Difference	AVR success rate
	☒ Positive, sign of infection	37	4	4	100 %	0 %	100 %
	Uncertain result	1	-				0 %
	Negative, no sign of infection	1	-				0 %
	Longstanding infection, serological scar	7	-				0 %
	Laboratory does not give clinical interpretations	28	-				-
	Total:	74	4	4	100 %	19.6 %	80.4 %


Sample S002 BopeAbA



Sample S002 BopeAbG



Sample S002 BopeAbM



BopeAbA	Interpretation	Method	Interpretation count	Method count	Own score	Max score	Own success rate	Difference	AVR success rate
	Negative		6		-				0 %
		Euroimmun Anti- B. pertussis Toxin ELISA IgA		2					
		Euroimmun Euroline Bordetella pertussis IgA		1					
		Orgentec B. pertussis TOXIN IgA ELISA		2					
		Virion/Serion B. pertussis IgA		1					
	Borderline		3		2	2	100 %	0 %	100 %
		Euroimmun Euroline Bordetella pertussis IgA		1					
		TestLine EIA Bordetella pertussis Toxin IgA		2					
	Weak positive		2		-				100 %
		Euroimmun Anti- B. pertussis Toxin ELISA IgA		1					
		Euroimmun Anti-Bordetella FHA ELISA IgA		1					
	Positive		37		-				100 %
		Diasorin LIAISON® Bordetella pertussis Toxin IgA		12					
		Euroimmun Anti- B. pertussis Toxin ELISA IgA		9					
		Euroimmun Anti-Bordetella FHA ELISA IgA		2					
		Euroimmun Euroline Bordetella pertussis IgA		1					
		Novatec NovaLisa B. pertussis IgA ELISA		1					
		Orgentec B. pertussis TOXIN IgA ELISA		10					
		TestLine BLOT-LINE Bordetella IgA		1					
		TestLine Microblot-Array Bordetella IgA		1					
	Uncertain		1		-				0 %
		Euroimmun Anti- B. pertussis Toxin ELISA IgA		1					
	Total:		49		2	2	100 %	14.3 %	85.7 %

BopeAbG	Interpretation	Method	Interpretation count	Method count	Own score	Max score	Own success rate	Difference	AVR success rate
	Negative		1		-				0 %
		Euroimmun Euroline Bordetella pertussis IgG		1					
	Borderline		3		-				0 %
		Diasorin LIAISON® Bordetella pertussis Toxin IgG		1					
		Euroimmun Anti-B. pertussis Toxin ELISA IgG		1					
		Euroimmun Anti-Bordetella FHA ELISA IgG		1					


	☒ Positive		70		2	2	100 %	0 %	100 %
		Diasorin LIAISON® Bordetella pertussis Toxin IgG		24					
		Euroimmun Anti-B. pertussis Toxin ELISA IgG		15					
		Euroimmun Anti-Bordetella FHA ELISA IgG		3					
		Euroimmun Euroline Bordetella pertussis IgG		2					
		Novagnost Bordetella pertussis toxin (PT) IgG		1					
		Novatec NovaLisa B. pertussis IgG ELISA		2					
		Orgentec B. pertussis TOXIN IgG ELISA		12					
		TestLine BLOT-LINE Bordetella IgG		1					
		☒ TestLine EIA Bordetella pertussis Toxin IgG		2					
		TestLine Microblot-Array Bordetella IgG		1					
		Vircell VIRCLIA Monotest Bordetella Pertussis Toxin IgG		4					
		Virion/Serion B. pertussis IgG		1					
		Virion/Serion B. pertussis Toxin IgG		2					
	Total:		74		2	2	100 %	5.4 %	94.6 %

BopeAbM	Interpretation	Method	Interpretation count	Method count	Own score	Max score	Own success rate	Difference	AVR success rate
	☒ Negative		11		2	2	100 %	0 %	100 %
		Euroimmun Anti-B. pertussis ELISA IgM		9					
		☒ TestLine EIA Bordetella pertussis Toxin IgM		2					
	Borderline		1		-				0 %
		Novatec NovaLisa B. pertussis IgM ELISA		1					
	Total:		12		2	2	100 %	8.3 %	91.7 %

Client report

	No of participants	No of responded participants	Response percentage
Bordetella pertussis, antibodies, April, 2-2025	77	70	90.9 %

Summary Result 2



Summary	Own score	Max score	Own success rate	Difference	AVR success rate
Sample S001	8	8	100 %	5.7 %	94.3 %
Sample S002	8	8	100 %	13.2 %	86.8 %
Average:			100 %	9.5 %	90.5 %


History	Test nr.	Own success rate	Difference	AVR success rate
Round 2024-4	1-1	87.5 %	0.3 %	87.2 %
Round 2024-4	1-2	87.5 %	0.3 %	87.2 %
Round 2024-4	1-3	87.5 %	0.3 %	87.2 %
Round 2024-2	1-1	100 %	10.6 %	89.4 %
Round 2024-2	1-2	100 %	10.6 %	89.4 %
Round 2024-2	1-3	100 %	10.6 %	89.4 %
Round 2023-4	1-1	60 %	-25.2 %	85.2 %
Round 2023-4	1-2	60 %	-25.2 %	85.2 %

Sample S001 Result 2




Sample S001 results	Responded	Own score	Max score	Own success rate	Difference	AVR success rate	Count
	Clinical interpretation	4	4	100 %	10.9 %	89.1 %	74
	BopeAbA	2	2	100 %	6.1 %	93.9 %	49
	BopeAbG	2	2	100 %	0 %	100 %	74
	BopeAbM	-	-	-	-	100 %	12
Total:		8	8	100 %	5.7 %	94.3 %	209

Sample S001 Clinical interpretation




Clinical interpretation	Interpretation	Interpretation count	Own score	Max score	Own success rate	Difference	AVR success rate
	☒ Positive, sign of infection	41	4	4	100 %	0 %	100 %
	Longstanding infection, serological scar	5	-				0 %
	Laboratory does not give clinical interpretations	28	-				-
	Total:	74	4	4	100 %	10.9 %	89.1 %


Sample S001 BopeAbA



Sample S001 BopeAbG



Sample S001 BopeAbM



BopeAbA	Interpretation	Method	Interpretation count	Method count	Own score	Max score	Own success rate	Difference	AVR success rate
	Negative		2		-				0 %
		Euroimmun Anti- B. pertussis Toxin ELISA IgA		1					
		Euroimmun Euroline Bordetella pertussis IgA		1					
	Positive		46		2	2	100 %	0 %	100 %
		Diasorin LIAISON® Bordetella pertussis Toxin IgA		12					
		Euroimmun Anti- B. pertussis Toxin ELISA IgA		11					
		Euroimmun Anti-Bordetella FHA ELISA IgA		3					
		Euroimmun Euroline Bordetella pertussis IgA		2					
		Novatec NovaLisa B. pertussis IgA ELISA		1					
		Orgentec B. pertussis TOXIN IgA ELISA		12					
		TestLine BLOT-LINE Bordetella IgA		1					
		TestLine EIA Bordetella pertussis Toxin IgA		2					
		TestLine Microblot-Array Bordetella IgA		1					
		Virion/Serion B. pertussis Toxin IgA		1					
	Uncertain		1		-				0 %
		Euroimmun Anti- B. pertussis Toxin ELISA IgA		1					
	Total:		49		2	2	100 %	6.1 %	93.9 %

BopeAbG	Interpretation	Method	Interpretation count	Method count	Own score	Max score	Own success rate	Difference	AVR success rate
	Positive		74		2	2	100 %	0 %	100 %
		Diasorin LIAISON® Bordetella pertussis Toxin IgG		25					
		Euroimmun Anti-B. pertussis Toxin ELISA IgG		17					
		Euroimmun Anti-Bordetella FHA ELISA IgG		3					
		Euroimmun Euroline Bordetella pertussis IgG		3					
		Novagnost Bordetella pertussis toxin (PT) IgG		1					
		Novatec NovaLisa B. pertussis IgG ELISA		2					
		Orgentec B. pertussis TOXIN IgG ELISA		12					
		TestLine BLOT-LINE Bordetella IgG		1					
		TestLine EIA Bordetella pertussis Toxin IgG		2					
		TestLine Microblot-Array Bordetella IgG		1					

		Vircell VIRCLIA Monotest Bordetella Pertussis Toxin IgG		4					
		Virion/Serion B. pertussis Toxin IgG		3					
	Total:		74		2	2	100 %	0 %	100 %


BopeAbM	Interpretation	Method	Interpretation count	Method count	Own score	Max score	Own success rate	Difference	AVR success rate
	Negative		12		-				100 %
		Euroimmun Anti-B. pertussis ELISA IgM		9					
		Novatec NovaLisa B. pertussis IgM ELISA		1					
		TestLine EIA Bordetella pertussis Toxin IgM		2					
	Total:		12		-	-	-	-	100 %

Sample S002 Result 2




Sample S002 results	Responded	Own score	Max score	Own success rate	Difference	AVR success rate	Count
	Clinical interpretation	4	4	100 %	19.6 %	80.4 %	74
	BopeAbA	2	2	100 %	14.3 %	85.7 %	49
	BopeAbG	2	2	100 %	5.4 %	94.6 %	74
	BopeAbM	-	-	-	-	91.7 %	12
Total:		8	8	100 %	13.2 %	86.8 %	209

Sample S002 Clinical interpretation




Clinical interpretation	Interpretation	Interpretation count	Own score	Max score	Own success rate	Difference	AVR success rate
	☒ Positive, sign of infection	37	4	4	100 %	0 %	100 %
	Uncertain result	1	-				0 %
	Negative, no sign of infection	1	-				0 %
	Longstanding infection, serological scar	7	-				0 %
	Laboratory does not give clinical interpretations	28	-				-
	Total:	74	4	4	100 %	19.6 %	80.4 %


Sample S002 BopeAbA



Sample S002 BopeAbG



Sample S002 BopeAbM



BopeAbA	Interpretation	Method	Interpretation count	Method count	Own score	Max score	Own success rate	Difference	AVR success rate
	Negative		6		-				0 %
		Euroimmun Anti- B. pertussis Toxin ELISA IgA		2					
		Euroimmun Euroline Bordetella pertussis IgA		1					
		Orgentec B. pertussis TOXIN IgA ELISA		2					
		Virion/Serion B. pertussis IgA		1					
	Borderline		3		-				100 %
		Euroimmun Euroline Bordetella pertussis IgA		1					
		TestLine EIA Bordetella pertussis Toxin IgA		2					
	Weak positive		2		-				100 %
		Euroimmun Anti- B. pertussis Toxin ELISA IgA		1					
		Euroimmun Anti-Bordetella FHA ELISA IgA		1					
	Positive		37		2	2	100 %	0 %	100 %
		Diasorin LIAISON® Bordetella pertussis Toxin IgA		12					
		Euroimmun Anti- B. pertussis Toxin ELISA IgA		9					
		Euroimmun Anti-Bordetella FHA ELISA IgA		2					
		Euroimmun Euroline Bordetella pertussis IgA		1					
		Novatec NovaLisa B. pertussis IgA ELISA		1					
		Orgentec B. pertussis TOXIN IgA ELISA		10					
		TestLine BLOT-LINE Bordetella IgA		1					
		TestLine Microblot-Array Bordetella IgA		1					
	Uncertain		1		-				0 %
		Euroimmun Anti- B. pertussis Toxin ELISA IgA		1					
	Total:		49		2	2	100 %	14.3 %	85.7 %

BopeAbG	Interpretation	Method	Interpretation count	Method count	Own score	Max score	Own success rate	Difference	AVR success rate
	Negative		1		-				0 %
		Euroimmun Euroline Bordetella pertussis IgG		1					
	Borderline		3		-				0 %
		Diasorin LIAISON® Bordetella pertussis Toxin IgG		1					
		Euroimmun Anti-B. pertussis Toxin ELISA IgG		1					


		Euroimmun Anti-Bordetella FHA ELISA IgG		1					
	🕒 Positive		70		2	2	100 %	0 %	100 %
		Diasorin LIAISON® Bordetella pertussis Toxin IgG		24					
		Euroimmun Anti-B. pertussis Toxin ELISA IgG		15					
		Euroimmun Anti-Bordetella FHA ELISA IgG		3					
		Euroimmun Euroline Bordetella pertussis IgG		2					
		Novagnost Bordetella pertussis toxin (PT) IgG		1					
		Novatec NovaLisa B. pertussis IgG ELISA		2					
		Orgentec B. pertussis TOXIN IgG ELISA		12					
		🕒 TestLine BLOT-LINE Bordetella IgG		1					
		TestLine EIA Bordetella pertussis Toxin IgG		2					
		TestLine Microblot-Array Bordetella IgG		1					
		Vircell VIRCLIA Monotest Bordetella Pertussis Toxin IgG		4					
		Virion/Serion B. pertussis IgG		1					
		Virion/Serion B. pertussis Toxin IgG		2					
	Total:		74		2	2	100 %	5.4 %	94.6 %

BopeAbM	Interpretation	Method	Interpretation count	Method count	Own score	Max score	Own success rate	Difference	AVR success rate
	Negative		11		-				100 %
		Euroimmun Anti-B. pertussis ELISA IgM		9					
		TestLine EIA Bordetella pertussis Toxin IgM		2					
	Borderline		1		-				0 %
		Novatec NovaLisa B. pertussis IgM ELISA		1					
	Total:		12		-	-	-	-	91.7 %

Client report

	No of participants	No of responded participants	Response percentage
Bordetella pertussis, antibodies, April, 2-2025	77	70	90.9 %

Summary Result 3



Summary	Own score	Max score	Own success rate	Difference	AVR success rate
Sample S001	8	8	100 %	5.7 %	94.3 %
Sample S002	8	8	100 %	13.2 %	86.8 %
Average:			100 %	9.5 %	90.5 %


History	Test nr.	Own success rate	Difference	AVR success rate
Round 2024-4	1-1	87.5 %	0.3 %	87.2 %
Round 2024-4	1-2	87.5 %	0.3 %	87.2 %
Round 2024-4	1-3	87.5 %	0.3 %	87.2 %
Round 2024-2	1-1	100 %	10.6 %	89.4 %
Round 2024-2	1-2	100 %	10.6 %	89.4 %
Round 2024-2	1-3	100 %	10.6 %	89.4 %
Round 2023-4	1-1	60 %	-25.2 %	85.2 %
Round 2023-4	1-2	60 %	-25.2 %	85.2 %

Sample S001 Result 3




Sample S001 results	Responded	Own score	Max score	Own success rate	Difference	AVR success rate	Count
	Clinical interpretation	4	4	100 %	10.9 %	89.1 %	74
	BopeAbA	2	2	100 %	6.1 %	93.9 %	49
	BopeAbG	2	2	100 %	0 %	100 %	74
	BopeAbM	-	-	-	-	100 %	12
Total:		8	8	100 %	5.7 %	94.3 %	209

Sample S001 Clinical interpretation




Clinical interpretation	Interpretation	Interpretation count	Own score	Max score	Own success rate	Difference	AVR success rate
	☒ Positive, sign of infection	41	4	4	100 %	0 %	100 %
	Longstanding infection, serological scar	5	-				0 %
	Laboratory does not give clinical interpretations	28	-				-
	Total:	74	4	4	100 %	10.9 %	89.1 %


Sample S001 BopeAbA



Sample S001 BopeAbG



Sample S001 BopeAbM



BopeAbA	Interpretation	Method	Interpretation count	Method count	Own score	Max score	Own success rate	Difference	AVR success rate
	Negative		2		-				0 %
		Euroimmun Anti- B. pertussis Toxin ELISA IgA		1					
		Euroimmun Euroline Bordetella pertussis IgA		1					
	☒ Positive		46		2	2	100 %	0 %	100 %
		Diasorin LIAISON® Bordetella pertussis Toxin IgA		12					
		Euroimmun Anti- B. pertussis Toxin ELISA IgA		11					
		Euroimmun Anti-Bordetella FHA ELISA IgA		3					
		Euroimmun Euroline Bordetella pertussis IgA		2					
		Novatec NovaLisa B. pertussis IgA ELISA		1					
		Orgentec B. pertussis TOXIN IgA ELISA		12					
		TestLine BLOT-LINE Bordetella IgA		1					
		TestLine EIA Bordetella pertussis Toxin IgA		2					
		☒ TestLine Microblot-Array Bordetella IgA		1					
		Virion/Serion B. pertussis Toxin IgA		1					
	Uncertain		1		-				0 %
		Euroimmun Anti- B. pertussis Toxin ELISA IgA		1					
	Total:		49		2	2	100 %	6.1 %	93.9 %

BopeAbG	Interpretation	Method	Interpretation count	Method count	Own score	Max score	Own success rate	Difference	AVR success rate
	☒ Positive		74		2	2	100 %	0 %	100 %
		Diasorin LIAISON® Bordetella pertussis Toxin IgG		25					
		Euroimmun Anti-B. pertussis Toxin ELISA IgG		17					
		Euroimmun Anti-Bordetella FHA ELISA IgG		3					
		Euroimmun Euroline Bordetella pertussis IgG		3					
		Novagnost Bordetella pertussis toxin (PT) IgG		1					
		Novatec NovaLisa B. pertussis IgG ELISA		2					
		Orgentec B. pertussis TOXIN IgG ELISA		12					
		TestLine BLOT-LINE Bordetella IgG		1					
		TestLine EIA Bordetella pertussis Toxin IgG		2					
		☒ TestLine Microblot-Array Bordetella IgG		1					

		Vircell VIRCLIA Monotest Bordetella Pertussis Toxin IgG		4					
		Virion/Serion B. pertussis Toxin IgG		3					
	Total:		74		2	2	100 %	0 %	100 %


BopeAbM	Interpretation	Method	Interpretation count	Method count	Own score	Max score	Own success rate	Difference	AVR success rate
	Negative		12		-				100 %
		Euroimmun Anti-B. pertussis ELISA IgM		9					
		Novatec NovaLisa B. pertussis IgM ELISA		1					
		TestLine EIA Bordetella pertussis Toxin IgM		2					
	Total:		12		-	-	-	-	100 %

Sample S002 Result 3




Sample S002 results	Responded	Own score	Max score	Own success rate	Difference	AVR success rate	Count
	Clinical interpretation	4	4	100 %	19.6 %	80.4 %	74
	BopeAbA	2	2	100 %	14.3 %	85.7 %	49
	BopeAbG	2	2	100 %	5.4 %	94.6 %	74
	BopeAbM	-	-	-	-	91.7 %	12
Total:		8	8	100 %	13.2 %	86.8 %	209

Sample S002 Clinical interpretation




Clinical interpretation	Interpretation	Interpretation count	Own score	Max score	Own success rate	Difference	AVR success rate
	☒ Positive, sign of infection	37	4	4	100 %	0 %	100 %
	Uncertain result	1	-				0 %
	Negative, no sign of infection	1	-				0 %
	Longstanding infection, serological scar	7	-				0 %
	Laboratory does not give clinical interpretations	28	-				-
	Total:	74	4	4	100 %	19.6 %	80.4 %


Sample S002 BopeAbA



Sample S002 BopeAbG



Sample S002 BopeAbM



BopeAbA	Interpretation	Method	Interpretation count	Method count	Own score	Max score	Own success rate	Difference	AVR success rate
	Negative		6		-				0 %
		Euroimmun Anti- B. pertussis Toxin ELISA IgA		2					
		Euroimmun Euroline Bordetella pertussis IgA		1					
		Orgentec B. pertussis TOXIN IgA ELISA		2					
		Virion/Serion B. pertussis IgA		1					
	Borderline		3		-				100 %
		Euroimmun Euroline Bordetella pertussis IgA		1					
		TestLine EIA Bordetella pertussis Toxin IgA		2					
	Weak positive		2		-				100 %
		Euroimmun Anti- B. pertussis Toxin ELISA IgA		1					
		Euroimmun Anti-Bordetella FHA ELISA IgA		1					
	Positive		37		2	2	100 %	0 %	100 %
		Diasorin LIAISON® Bordetella pertussis Toxin IgA		12					
		Euroimmun Anti- B. pertussis Toxin ELISA IgA		9					
		Euroimmun Anti-Bordetella FHA ELISA IgA		2					
		Euroimmun Euroline Bordetella pertussis IgA		1					
		Novatec NovaLisa B. pertussis IgA ELISA		1					
		Orgentec B. pertussis TOXIN IgA ELISA		10					
		TestLine BLOT-LINE Bordetella IgA		1					
		TestLine Microblot-Array Bordetella IgA		1					
	Uncertain		1		-				0 %
		Euroimmun Anti- B. pertussis Toxin ELISA IgA		1					
	Total:		49		2	2	100 %	14.3 %	85.7 %

BopeAbG	Interpretation	Method	Interpretation count	Method count	Own score	Max score	Own success rate	Difference	AVR success rate
	Negative		1		-				0 %
		Euroimmun Euroline Bordetella pertussis IgG		1					
	Borderline		3		-				0 %
		Diasorin LIAISON® Bordetella pertussis Toxin IgG		1					
		Euroimmun Anti-B. pertussis Toxin ELISA IgG		1					
		Euroimmun Anti-Bordetella FHA ELISA IgG		1					

	☒ Positive		70		2	2	100 %	0 %	100 %
		Diasorin LIAISON® Bordetella pertussis Toxin IgG		24					
		Euroimmun Anti-B. pertussis Toxin ELISA IgG		15					
		Euroimmun Anti-Bordetella FHA ELISA IgG		3					
		Euroimmun Euroline Bordetella pertussis IgG		2					
		Novagnost Bordetella pertussis toxin (PT) IgG		1					
		Novatec NovaLisa B. pertussis IgG ELISA		2					
		Orgentec B. pertussis TOXIN IgG ELISA		12					
		TestLine BLOT-LINE Bordetella IgG		1					
		TestLine EIA Bordetella pertussis Toxin IgG		2					
		☒ TestLine Microblot-Array Bordetella IgG		1					
		Vircell VIRCLIA Monotest Bordetella Pertussis Toxin IgG		4					
		Virion/Serion B. pertussis IgG		1					
		Virion/Serion B. pertussis Toxin IgG		2					
	Total:		74		2	2	100 %	5.4 %	94.6 %

BopeAbM	Interpretation	Method	Interpretation count	Method count	Own score	Max score	Own success rate	Difference	AVR success rate
	Negative		11		-				100 %
		Euroimmun Anti-B. pertussis ELISA IgM		9					
		TestLine EIA Bordetella pertussis Toxin IgM		2					
	Borderline		1		-				0 %
		Novatec NovaLisa B. pertussis IgM ELISA		1					
	Total:		12		-	-	-	-	91.7 %