



ORIGINAL EVALUATION

VR3-A 2025 Infectious Disease Serology

INSTITUTION: TestLine Clinical Diagnostics SRO
Brno CZ 61200

ATTENTION: Lenka Hanakova

CAP NUMBER: 9670036-01 **Kit#** 1

KIT INFORMATION:	Kit ID:	Kit Mailed:	Original Evaluation:	Next Mailing Date:
	38733081	3/12/2025	4/18/2025	10/27/2025

COPIED TO: CAP

LEGEND: **Exception Reason Codes appearing in this evaluation:**
[26] = Educational challenge.
[30] = Scientific committee decision.

Reviewed By

Date



**COLLEGE of AMERICAN
PATHOLOGISTS**

325 Waukegan Road, Northfield, Illinois 60093-2750
800-323-4040 • cap.org

CAP Number: 9670036-01

Kit# 1

Institution: TestLine Clinical Diagnostics SRO

Attention: Lenka Hanakova

City / State: Brno CZ 61200

Kit ID: 38733081

Kit Mailed: 3/12/2025

Original Evaluation: 4/18/2025

EVALUATION

ORIGINAL

VR3-A 2025 Infectious Disease Serology

Test Method	Specimen	Your Result	Intended Response	Your Grade
CMV, IgG OTHER	VR3-01	PRESENT	PRESENT	Good
CMV, IgM OTHER	VR3-01	ABSENT	ABSENT	Good
EBV, VCA, IgG OTHER	VR3-02	ABSENT		See Note [26]
EBV, VCA, IgM OTHER	VR3-02	ABSENT		See Note [26]
EBV, EA, IgG OTHER	VR3-02	ABSENT		See Note [26]
EBV, EBNA, IgG OTHER	VR3-02	ABSENT		See Note [26]
EBV, EBNA, IgM OTHER	VR3-02	ABSENT		See Note [26]
H pylori, IgG OTHER	VR3-03	ABSENT	ABSENT	Good
H pylori, IgA OTHER	VR3-03	ABSENT	ABSENT	Good

The College of American Pathologists recommends that the result of this interlaboratory comparison not be used as a sole criterion for judging the performance of any individual clinical laboratory.



**COLLEGE of AMERICAN
PATHOLOGISTS**

325 Waukegan Road, Northfield, Illinois 60093-2750
800-323-4040 • cap.org

CAP Number: 9670036-01

Kit# 1

Institution: TestLine Clinical Diagnostics SRO

Attention: Lenka Hanakova

City / State: Brno CZ 61200

Kit ID: 38733081

Kit Mailed: 3/12/2025

Original Evaluation: 4/18/2025

EVALUATION

ORIGINAL

VR3-A 2025 Infectious Disease Serology

Test Method	Specimen	Your Result	Intended Response	Your Grade
HSV 1, IgG OTHER	VR3-04	PRESENT		See Note [30]
HSV 2, IgG OTHER	VR3-04	PRESENT	PRESENT	Good
HSV, NOS, IgG OTHER	VR3-04	PRESENT	PRESENT	Good
M pneu, IgG OTHER	VR3-05	ABSENT		See Note [26]
M pneu, IgM OTHER	VR3-05	ABSENT		See Note [26]
Rubeola, IgG OTHER	VR3-06	ABSENT	PRESENT	Unacceptable
Toxo, IgG OTHER	VR3-07	ABSENT	ABSENT	Good
Toxo, IgM OTHER	VR3-07	ABSENT	ABSENT	Good
VZV, IgG OTHER	VR3-08	ABSENT	ABSENT	Good

The College of American Pathologists recommends that the result of this interlaboratory comparison not be used as a sole criterion for judging the performance of any individual clinical laboratory.