

SEKK s.r.o., EQA division, Arnošta z Pardubic 2605, 530 02 Pardubice, Czech Republic
Proficiency testing provider no. 7004 accredited by CAI according to ISO/IEC 17043.

Details concerning the subject and scope of the accreditation should be found on: <https://www.sekk.cz>

RESULT SHEET

EQA round: CRP2/25 - C-Reactive Protein

Deadline (EQA round closed): 2.5.2025

Professional supervision: Czech Society of Clinical Biochemistry (CSCB)

Participant: TestLine Clinical Diagnostics s.r.o.

Křižíkova 188/68, 612 00 Brno, Czech Republic CID: 47913240

Bc. Lenka Hanáková

Key	PR = participant's result:	Quantitative results
C = overall assessment of the test:	✓ ... correct result	AV = assigned value
+ ... successful test	✗ ... wrong result	Dmax = acceptable difference
- ... unsuccessful test	• ... not evaluated	LL = lower limit
± ... not evaluated	D[%] = percent difference (PR and AV)	UL = upper limit

Test	[unit]	C	Participant's basic information about test						Success in 2 years
Sample			Comparability						
Evaluated group			AV	Dmax	LL	UL	PR	D[%]	
C-reactive protein	[mg/L]	+	M=99=Another measurement principle, R=140=TestLine						75 %
Sample A									
All results			25,4	24 %	19,3	31,5	23,78	-6,4 ✓	
Sample B									
All results			82,4	24 %	62,6	103	67,25	-18 ✓	

Your total success in the rounds of this programme during last 2 years:

2023: CRP4 = 100 %

2024: CRP2 = 0 % CRP4 = 100 %

2025: CRP2 = 100 %

Your total success in this programme during last 2 years: 75 % (number of evaluated tests: 4)

Summary evaluation of this round should be found on: <https://www.sekk.cz>