

## EIA Adenovirus IgA



**EAN Code:** 8595635308840

**Catalog number:** AdA096

**Package size:** 96 tests

**Storage:** 2-8 °C

**Producer:** TestLine Clinical Diagnostics s.r.o.

### Description:

- A native antigen is bound to the well surface of the microtiter plate.
- If specific antibodies are present, they bind to the antigen, are labeled by the Conjugate in the following steps and are detected by color reaction with a single component substrate (TMB-Complete).
- The kit allows 96 tests, including controls in a split microtiter plate with color-coded strips and breakable wells.

### Advantages:

- The total assay time is about 2 hours.
- High sensitivity and specificity of the test.
- Kit includes CUT-OFF, Positive Control and Negative Control.
- Semi-quantitative evaluation in the Index of Positivity (IP).
- Ready-to-use, color-coded components.
- Single-component substrate.
- Interchangeable components with the exception of kit specific components (Controls, Conjugate, Plate).

### Application:

- Screening test for Adenovirus infection.

### Brief assay procedure:

1. Dilute samples (1:101).

2. Incubation at 37 °C for 10 min.
3. Pipette Controls and diluted Samples.
4. Incubate at 37°C for 30 min.
5. Aspirate and wash the wells 5x
6. Pipette Conjugate.
7. Incubate at 37°C for 30 min.
8. Aspirate and wash the wells 5x
9. Pipette Substrate (TMB-Complete).
10. Incubate at 37°C for 30 min.
11. Pipette Stop Solution.
12. Read color intensity at 450 nm.
13. Evaluate the test.