

EIA VZV IgM



EAN Code: 8595635304248

Catalog number: VZVM96

Package size: 96 tests

Storage: 2-8 °C

Producer: TestLine Clinical Diagnostics s.r.o.

Description:

- Microtitre wells are coated with the purified and inactivated VZV antigen with high content of specific immuno dominant epitopes.
- If specific antibodies are present, they bind to the antigen, are labeled by the Conjugate in the following steps and are detected by color reaction with a single component substrate (TMB-Complete).
- The kit allows 96 tests, including controls in a split microtiter plate with color-coded strips and breakable wells.

Advantages:

- The total assay time is about 2 hours.
- High sensitivity and specificity of the test.
- Kit includes CUT-OFF, Positive Control and Negative Control, Calibrators.
- Semi-quantitative evaluation in the Index of Positivity (IP).
- Ready-to-use, color-coded components.
- Single-component substrate.

- Interchangeable components with the exception of kit specific components (Controls, Conjugate, Plate).
- Sample diluent contains IgG/RF sorbent.

Application:

- Screening test for the detection of specific IgM antibodies in human serum or plasma.
- Disease phase diagnostics, IgM antibodies are indicator of active infection.

Brief assay procedure:

1. Dilute samples (1:101), cerebrospinal fluid (1:3).
2. Pipette Controls and diluted Samples.
3. Incubate at 37°C for 30 min.
4. Aspirate and wash the wells 5x
5. Pipette Conjugate.
6. Incubate at 37°C for 30 min.
7. Aspirate and wash the wells 5x
8. Pipette Substrate (TMB-Complete).
9. Incubate at 37°C for 30 min.
10. Pipette Stop Solution.
11. Read color intensity at 450 nm.
12. Evaluate the test.