

Microblot-Array Bordetella IgM

EAN Code: 8595635310638

Catalog number: BpMMA48

Package size: 48 tests

Storage: 2-8 °C

Producer: TestLine Clinical Diagnostics s.r.o.



Description:

- Used Antigen: combination of selected parts of the specific antigens of Bordetella pertussis (PT, FHA, ACT, TCF), and Bordetella parapertussis (Pertactin, FimN EntA).
- Recombinant antigens are applied in the form of microdots (spots) in triplets to a nitrocellulose membrane in a microtiter well format.
- If specific antibodies are present in the sample, they will bind to the respective antigens.
- The complex is labelled with Conjugate and detected through a colour reaction with substrate (BCIP/NBT).
- The kit enables 48 tests.
- Short incubation periods, total assay time: approximately 1.5 h.
- Each individual test includes control spots, which serve to check functionality and sensitivity of the sets, and also the spots for checking the presence of the conjugate.
- For quantitative evaluation the wells contain calibration spots.
- All reagents supplied are ready to use.
- The kit may be used sequentially for smaller batches of samples (reagents are provided in sufficient quantities).
- Evaluation using Microblot-Array reader and the appropriate software.

Applications:

- Differential diagnostics of Bordetella pertussis and parapertussis.
- Confirmatory method for ELISA and/or agglutination reaction.

Brief assay procedure:

- Dilute samples serum/plasma (1:51).
- Pipette diluted samples.

3. Incubate at room temperature for 30 min.
4. Wash 3 times for 5 min.
5. Add Conjugate.
6. Incubate at room temperature for 30 min.
7. Wash 3 times for 5 min.
8. Add Substrate (BCIP/NPT).
9. Incubate at room temperature for 15 min.
10. Wash 2 times for 5 min in distilled water.
11. Dry and evaluate of test.