

## **Microblot-Array EBV IgM**



EAN Code: 8595635308765 Catalog number: EBMMA96 Package size: 96 tests

Storage: 2-8 °C

Producer: TestLine Clinical Diagnostics s.r.o.

## **Description:**

- Used Antigens: combination of selected parts of the specific antigens of EBV (EBNA-1, EBNA-2, VCA p18, VCA p23, EA-D p54, EA-D p138, EA-R, Rta, ZEBRA, gp85, gp350 and LMP1).
- Recombinant antigens are applied in the form of microdots (spots) in triplets to a nitrocellulose membrane in a microtiter well format.
- If specific antibodies are present in the sample, they will bind to the respective antigens.
- The complex is labelled with Conjugate and detected through a colour reaction with substrate (BCIP/NBT).
- The kit enables 96 tests.
- Short incubation periods, total assay time: approximately 1.5 h.
- Each individual test includes control spots, which serves to check functionality and sensitivity of the sets, and also the spots for checking the presence of the conjugate.
- Positive controls included enabling validation of the tests.
- For quantitative evaluation the wells contain calibration spots.
- All reagents supplied are ready to use.



- The kit may be used sequentially for smaller batches of samples (reagents are provided in sufficient quantities).
- Evaluation using Microblot-Array reader and the appropriate software.

## **Applications:**

- Laboratory test for the detection of EBV
- Confirmatory test to ELISA.

## **Brief assay procedure:**

- 1. Dilute samples serum/plasma (1:51).
- 2. Pipette diluted samples.
- 3. Incubate at room temperature for 30 min.
- 4. Wash 3 times for 5 min.
- 5. Add Conjugate.
- 6. Incubate at room temperature for 30 min.
- 7. Wash 3 times for 5 min.
- 8. Add Substrate (BCIP/NPT).
- 9. Incubate at room temperature for 15 min.
- 10. Wash 2 times for 5 min in distilled water.
- 11. Dry and evaluate of test.

Exportováno 18. 10. 2025 Stránka 2 / 2