


Microblot Array Reader



Producer: TestLine Clinical Diagnostics s.r.o.



Array Reader C-series

- Colorimetric reader optimized for 96-well microtiter plate format
- Image analysis and evaluation of tests based on the intensity of individual spots
- Scan speed approx. 5 minutes / full microtiter plate
- Possibility to scan selected wells or strips
- Automatic localization and capture of spots
- Quantitative evaluation in U/ml
- Assessment of test validity through control spots
- Clear display of results
- Two-way communication with the information system

TestLine Analytics Software

- The user-friendly evaluation program intuitively accompanies the entire processing and evaluation of the results, including communication with LIS
- The evaluation process is completely automatic
- Centring and image processing
- Image analysis and evaluation of tests based on the intensity of individual spots
- Assessment of test validity through control spots
- Clear display of results

Exportováno 5. 12. 2025 Stránka 1 / 1